Visual Inspection Log

Operations and Maintenance Plan

- Use for the Hg5 amalgam separator to ensure proper operation and maintenance plan.
- Visual inspection of separator once a month with date and signature.
- Collection container to be replaced when solids reach the full line or once a year; whichever comes first.

Office name: _____

Account number: _____

Date (MM/DD/YEAR)	Signature

Log must be kept current and maintained at a minimum of three years.

1.800.216.5505
www.solmetex.com